Paul Fulbrook - Editor

Lynne Harrison - Editor

The World of Critical Care Nursing

2009 Volume 7 Number 1

EDITORIAL

Paul Fulbrook, Professor of Nursing, National Centre for Clinical Outcomes Research, Australian Catholic University, and Nursing Director Research & Practice Development, The Prince Charles Hospital, Brisbane, Australia

BED SAFETY

Perhaps the most taken for granted piece of equipment that is used in modern health care is the hospital bed. However, in recent decades the 'humble' bed has morphed from being a basic piece of ironmongery to become a sophisticated therapeutic tool. Nowhere is this more apparent than in the critical care setting, where the bed is used not only for comfort and to position the patient effectively but often to deliver treatments such as cooling and warming, physiotherapy, and rotational therapy.

Although the hospital bed is a vital, it poses a risk to patients. In the critical care setting especially, patients are often sedated, confused and/or disorientated. Furthermore, their critical illness usually means that they are weak and, in addition, they often have muscle loss and joint stiffness. These problems may also be compounded by deteriorating sight and hearing, which are frequently found in the older population. These factors are particularly important to consider during the patient's recovery phase, when they are starting to become more mobile. At this time, they are more likely to be moved from the bed for short periods. For example, to sit in a chair as part of their rehabilitative process. As they gain strength they begin to move around in the bed more. This is a time of significant risk, when the potential for the patient to fall from the bed is greater. Vigilant nursing observation is required.

There is a growing body of evidence that highlights the risks associated with hospital beds: falls, entrapment, strangulation and even death. In this context, there are two critical factors: the height of the bed and the use of bed rails. It has been known for some time that 70-90% of falls, in the health care setting, occur as the patient tries to exit the bed for toileting purposes.

Historically, bed rails have been used as a safety measure to protect patients from falling from bed. Their use was perpetuated, in the USA especially, due to litigation payouts when patients sustained injuries from falls when bed rails were not used. However, current evidence indicates that although patients in low, rail-less beds may fall more often, they have no greater injury rate compared to patients in high beds using half-rails or no rails. 50-90% of falls from bed in hospital occur despite bed rails being applied (Oliver, 2002). Bed rails do not prevent falls; they simply increase the height of the fall and therefore increase the extent of the injury.

Contents

CONNECT en Español: reto, sueño o necesidad? [CONNECT in Spanish: challenge, dream or necessity?]

Wilson Cañón, Laura María Alberto.....89

Using monitoring technology in Cameroon

Gibson D'Cruz, Kate Allen.....92

Noise in the intensive care unit: a summary review

Gian Domenico Giusti, Federica Piergentili.....97

The role of the nurse in ventilator-associated pneumonia

Amy Kit Man Kwong, Sek Ying Chair, Lorna Kwai Ping

Suen.....100

Ethical dilemma: do not resuscitate a vegetative paediatric patient

Wong Pik Yan Pat, Lorna Kwai Ping Suen, Sek Ying Chair

.....103

Connect: The World of Critical Care Nursing is the official journal of the World Federation of Critical Care Nurses and is published in association with the European federation of Critical Care Nursing associations.

The risk of strangulation and suffocation can be reduced if correct bed rails are used. International Standard 60601-2-38 governs hospital beds. It states that bed rails should not have a gap between the rungs greater than 120mm and there should be a 60mm maximum gap between the mattress and the bed rail (FDA, 2006).

In essence, the way to reduce the risk of falls and injury is simple: lower the bed and do not use bed rails. If the latter are used, then half rails are better, as they reduce the likelihood that a patient will attempt to climb over them.

REFERENCES

Food and Drug Administration (2006). Hospital Bed System Dimensional and Assessment Guidance to Reduce Entrapment – Guidance for Industry and FDA Staff. USA: FDA.

Oliver D (2002). Bed falls and bedrails – what should we do? Age and Ageing 31(5), 415-418.

Editors

Paul Fulbrook (Australia), paul.fulbrook@acu.edu.au
Lynne Harrison (UK), lharrison2@uclan.ac.uk

Associate Editors

Birte Baktoft (Denmark), bba@rc.aaa.dk
Ged Williams (Australia), Ged.Williams@health.qld.gov.au
Karl Oyri (Norway), karl.oyri@deltadigital.no
Laura Alberta (Argentina), lauramalberty@yahoo.com
Patricia Moloney-Harmon (USA), pmooney@lifebridgehealth.org
Sek Ying Chair (Hong Kong), sychair@cuhk.edu.hk

International Advisory Boards

Africa

Juan Scribante (South Africa), scribantej@medicine.wits.ac.za

Asia-Pacific

CHAIR: Sek Ying Chair (Hong Kong), sychair@cuhk.edu.hk
David Chan (Hong Kong), dchan123@netvigator.com
Dongoak Debbie Kim (South Korea), dokim@djju.ac.kr

Eve Sison (Philippines), evesison2000@yahoo.com
Gordon Speed (New Zealand), speedg@wave.co.nz
Violeta Lopez (Australia), violeta.lopez@acu.edu.au

Europe

CHAIR: Birte Baktoft (Denmark), bba@rc.aaa.dk
Amparo Bernat Adell (Spain), a.bernat.a@telefonica.net
Anu Isohanni (Finland), anu.isohanni@hus.fi
Aysel Badir (Turkey), abadir@ku.edu.tr
Daniel Benlahoués (France), benla@chello.fr
Elio Drigo (Italy), edrigo@aniarti.it
Elizabeth Papathanassoglou (Greece), elipapa@nurs.uoa.gr
Drago Satošek (Slovenia), drago.satosek@kclj.si
Jos Latour (Netherlands), jos.latour@planet.nl
Lerma Hernandez (Belgium), lerma.hernandez@skynet.be
Marcel van Vilet (Netherlands), m.w.vanvilet@amc.uva.nl
Mona Ringdal (Sweden), mona.ringdal@fhs.qu.se
Rósa Þórsteinsdóttir (Iceland), rosa.th@simnet.is

Sabine Ruedebusch (Germany), Sabine.Ruedebusch@pius-hospital.de
Stefan Böhm (Austria), stefan.boehm@uklibk.ac.at
Tracy Pilcher (UK), tapilcher@aol.com

Latin America

CHAIR: Laura Alberto (Argentina), lauramalberty@yahoo.com
Berta Gamarra (Venezuela), sharo35@yahoo.com
Elina Siviero (Venezuela), elisiviero@hotmail.com
Mariana Mejias (Venezuela), mmjias@vc.edu.ve
Nelly Alconz (Bolivia), nelly_hazel@yahoo.com
Paola Maria Chauca (Perú), pchauca@hotmail.com
Patricia Morales (Perú), symbolo2001@yahoo.es
Wilson Cañón (Colombia), wilcamo32@yahoo.com

Layout

Jamie Fulbrook, jamie@jnewmedia.com

Publisher

Connect: The World of Critical Care Nursing is published by Paul Fulbrook, Jamie Fulbrook, and Lynne Harrison.

Disclaimer - All rights reserved; no part of this publication may be reproduced, electrical, mechanical, photocopying, recording, or otherwise without prior written permission from the publisher. The publishers, editor and editorial board wish to make it clear that the data, opinions and statements appearing in the articles herein are those of the contributor(s) concerned; such opinions are not necessarily shared by the editor or the editorial board. Accordingly, the publishers, editors and editorial board and their respective employees, officers and agents accept no liability for the consequences of any such inaccurate or misleading data, opinions or statements.

ISSN 1748-6254